

## EU Declaration of Conformity (DoC)

Tempress A/S hereby ensures and declares that

### Pressure gauges

Description of the pressure exposed element	Spiral wound stainless or steel pipe Bourdon stainless steel or bronze pipe
Drawing categories	Main group: 05
Maximal allowable operating pressure /+ overpressure	100% of full scale / +30%
The volume of the pressure exposed element do not exceed:	0,1 litre
Max/min allowable temperature exposure, std.	-25°C / +75°C
Fluida groupe:	1
Fluids:	Not applicable
Conformity assessment procedure followed	Article 3. pg. 3 and Module A pressure above 1000 bar
Other technical standards and specifications used: - Bourdon tube pressure gauges: - Requirements and test methods for pressure gauges and pressure switches:	DS/EN 837-1/AC  EN 12094-10

They are manufactured in accordance with the Directive 2014/68/EU of 15-05-2014 on the approximation of the laws of the Member States concerning pressure equipment.

Risskov, 07-03-2017



Søren Kruhøffer  
CEO

Document no.	Edition	Changing Message	Date
1007	2	Change of authority	07-03-2017
1007	1	Change of authority	06-11-2013